

Brucella OPS iELISA

Product #
B1110



THE WORLD'S MOST VERSATILE ELISA KIT FOR CATTLE BRUCELLOSIS

The Ellie's **BRUCELLA OPS iELISA** is a qualitative test kit based on indirect ELISA technology. Plates are coated with *B. abortus* O-chain polysaccharide (OPS), which is the same antigen used in the Brucella FPA (product code B1001).

THE FIRST ELISA THAT USES THE SAME ANTIGEN AS FPA

Antibodies against OPS are indicative of a recent or current infection with a smooth Brucella biovar.

RUNS INDIVIDUAL OR POOLED SERUM SAMPLES & INDIVIDUAL OR BULK MILK SAMPLES

This test can be used to monitor bulk milk antibody levels, for export/import purposes, for confirmatory testing or screening in vaccinated and non-vaccinated areas. It can also supplement the Brucella FPA in mass testing campaigns.

The Brucella OPS iELISA is a robust, versatile, highly sensitive, and highly specific test. It complements Brucella FPA by adding bulk milk testing and high sensitivity in testing animals in negative population.

B1110		Brucella Antibody Test Kit, iELISA OPS	
Trade Name	Brucella OPS iELISA		
Technology	Indirect ELISA		
Antigen	<i>B. abortus</i> & <i>B. melitensis</i> O-chain polysaccharide		
Analyte	IgG		
Species	Bovine		
Samples	Individual or pooled serum & individual or pooled milk		
Description	This test uses the same antigen as the Brucella FPA. Titers on OPS iELISA display a relative correlation to FPA results. This test is fully compliant with all applicable EU regulations (EU Council Directive 64/432/EEC, Annex C). It is recommended for eradication and control purposes, especially when milk samples are used.		

Samples	Sensitivity	Specificity	Reference method
Negative serum samples n=1454 Samples originated from several small farms across Serbia.	n.a.	99,6%	RBT, iELISA, cELISA
Positive serum samples n=204 Samples originated from the USDA/NVSL Serum Bank (United States of America). All were confirmed with culture.	100%.	n.a.	culture, CFT
Negative individual milk samples n=127 Samples originated from several small farms across Serbia (n=46) and dairy farms around Torreon, Mexico.	n.a.	99,2%	iELISA
Negative bulk milk samples n=189 Samples originated from single dairy processor in Serbia (n=46).	n.a.	100%	iELISA
Serum RB51 vaccinated animals n=20 Serum samples originated from the Larissa region in Greece and were provided by the National Reference Laboratory. No prior cases of field brucellosis had been reported in the herd. It is important to note that 40% of these samples were positive on some iELISA test kits and on the Rose Bengal test. They were negative on ELLIE OPS iELISA kit.	n.a.	100%	CFT, FPA

Standard	Sample Type	Dilution	Detection
OIEELISAspSS	Individual serum	1:16	Positive
OIEELISAspSS	Individual serum	1:64	Negative
OIEELISAspSS	Pooled serum samples	1:16 than 1:5	Positive
OIEELISAspSS	Pooled serum samples	1:64 than 1:5	Negative
OIEELISAspSS	Pooled milk samples	1:125 than 1:10	Positive
OIEELISAnSS	Pooled milk samples	1:10	Negative

ellie is the most trusted name in brucellosis diagnostics worldwide. With a full portfolio of test kits and reagents, we cover the entire field of brucellosis eradication in production animals. Our diagnostics are used as reference, confirmatory, and screening tools across the globe.

ellie

Headquarters

Ellie LLC

N114 W19320 Clinton Dr. #5

Germantown, WI 53022, U.S.A.

Phone: +1 (800) 556-6953 • ellie@ellielab.com

www.ellielab.com

Manufacturing

DOO Biotehnika IVD

Ratina 175, 36212 Kraljevo, Serbia

Phone: +381365861988, +381365861989

info@biotehnika.rs • www.ellielab.com